


RESTECH

For Trade Est.

Prop.: Eng. Ahmed Obeid

C.R.: 1010070990

C.R.: 7002028327



مؤسسة

البحاثة للتجارة

لصاحبها : المهندس أحمد عبید

سجل تجاري ١٠١٠٠٧٠٩٩٠

سجل تجاري ٧٠٠٢٠٢٨٣٢٧

CERTIFICATE OF CONFORMITY**Certificate#****S5004****Date: 27 AUGUST 2024**

System Sure Plus Hygiene Monitoring systems are manufactured to ISO 9000 Quality Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive 73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by 93/68 EEC.

SYSTEM SURE PLUS

Portable Luminometer

The Ensure Plus instrument with serial number **065676** has been tested and Performs in accordance with the product quality control specifications.

ATP Positive Control:

Positive Rod	Measured value(RLU)	Result
40-80	60	PASS

ATP Negative Control:

Negative Rod	Measured value(RLU)	Result
0 - 4	2	PASS

Engineer Signature:

**Valid to: 26 AUGUST 2025**

٠٠٦٩٩ (١١) ٤٦٤٣٧٧٠ فاكس،

٤٦٢١٦٣٣ / ٤٦٥٤٠٣٣ / ٤٦٥٩٥٩٧ / ٠٠٩٦٦ (١١) ٤٦٥٣٤٨٤ تلفون،

١١٤٧٢ رمز بريدي ٧٧٠ الرياض ص.ب

٠٠٦٩٩ (١٢) ٦٧٥٧٨٤٩ فاكس،

٦٧٥٧٩٦٤ / ٦٧٥٧٩٦٥ / ٠٠٩٦٦ (١٢) ٦٧٣٥١٩٢ تلفون،

٢١٤٥٣ رمز بريدي ١١٢٥٧ جدة ص.ب

٠٠٦٩٩ (١٣) ٨٦٧٠١٦١ فاكس،

٠٠٩٦٦ (١٣) ٨٦٧١٣١٦ تلفون،

الخبر:

Head Office: P.O.Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail : restech@restech.com.sa

Branches: P.O.Box 11257 Jeddah 21453 Phone : +966 (12) 6735192 / 6757964 / 6757965 Fax: +966 (12) 6757849 E-mail: jeddah.restech@restech.com.sa

Al-Khobar :

Phone : +966 (13) 8671316 Fax : +966 (13) 8670161

E-mail: dammam.restech@restech.com.sa

www.restech.com.sa