


RESTECH

For Trade Est.

Prop.: Eng. Ahmed Obeid

C.R.: 1010070990

C.R.: 7002028327



مؤسسة

البحاثة للتجارة

لصاحبها : المهندس أحمد عبيد

سجل تجاري ١٠١٠٧٠٩٩٠

سجل تجاري ٧٠٠٢٠٢٨٣٢٧

CERTIFICATE OF CONFORMITY**Certificate#****S5005**

Date: 15 OCTOBER 2024

System Sure Plus Hygiene Monitoring systems are manufactured to ISO 9000 Quality Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive 73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by 93/68 EEC.

System Sure Plus

Portable Luminometer

The S S Plus instrument with serial number **076841** has been tested and Performs in accordance with the product quality control specifications.

ATP Positive Control:

Positive Rod	Measured value(RLU)	Result
40-80	60	PASS

ATP Negative Control:

Negative Rod	Measured value(RLU)	Result
0 - 4	0	PASS

Engineer Signature:

Valid to: 14 OCT 2025

فاكس: ٤٦٤٣٧٧٠ (١١) ٠٠٦٩٩

فاكس: ٦٧٥٧٨٤٩ (١٢) ٠٠٦٩٩

فاكس: ٨٦٧٠١٦١ (١٣) ٠٠٦٩٩

تلفون: ٤٦٥٣٤٨٤ (١١) ٠٠٩٦٦ / ٤٦٥٤٠٣٣ / ٤٦٥٩٥٩٧

تلفون: ٦٧٣٥١٩٢ (١٢) ٠٠٩٦٦ / ٦٧٥٧٩٦٤ / ٦٧٥٧٩٦٥

تلفون: ٨٦٧١٣١٦ (١٣) ٠٠٩٦٦

المكتب الرئيسي: الرياض ص.ب ٧٧٧٠ رمز بريدي ١١٤٧٢

الفروع: جدة ص.ب ١١٢٥٧ رمز بريدي ٢١٤٥٣

الخبر:

Head Office: P.O.Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail : restech@restech.com.sa

Branches: P.O.Box 11257 Jeddah 21453 Phone : +966 (12) 6735192 / 6757964 / 6757965 Fax: +966 (12) 6757849 E-mail: jeddah.restech@restech.com.sa

Al-Khobar : Phone : +966 (13) 8671316 Fax : +966 (13) 8670161

E-mail: dammam.restech@restech.com.sa

www.restech.com.sa