


RESTECH**For Trade Est.****Prop.: Eng. Ahmed Obeid****C.R.: 1010070990****C.R.: 7002028327**

مؤسسة

البحاثة للتجارة

لصاحبها : المهندس أحمد عبيد

سجل تجاري ١٠١٠٠٧٠٩٩٠

سجل تجاري ٧٠٠٢٠٢٨٣٢٧

CERTIFICATE OF CONFORMITY**Certificate#****S5005****Date: 5 MAY 2025**

SSP 4 Hygiene Monitoring systems are manufactured to ISO 9000 Quality Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive 73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by 93/68 EEC.

System sure plus

Portable Luminometer


The SSP4 instrument with serial number **111671** has been tested and Performs in accordance with the product quality control specifications.

ATP Positive Control:

Positive Rod	Measured value(RLU)	Result
40-80	51	PASS

ATP Negative Control:

Negative Rod	Measured value(RLU)	Result
0 - 4	1	PASS

Engineer Signature:**Valid to: 4 MAY 2026**

فاكس: ٤٦٤٣٧٧٠ (١١) ٠٠٦٩٩

فاكس: ٦٧٥٧٨٤٩ (١٢) ٠٠٦٩٩

فاكس: ٨٦٧٠١٦١ (١٣) ٠٠٦٩٩

Head Office: P.O.Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail : restech@restech.com.sa
 Branches: P.O.Box 11257 Jeddah 21453 Phone : +966 (12) 6735192 / 6757964 / 6757965 Fax : +966 (12) 6757849 E-mail: jeddah.restech@restech.com.sa
 Al-Khobar : Phone : +966 (13) 8671316 Fax : +966 (13) 8670161 E-mail: dammam.restech@restech.com.sa
 www.restech.com.sa

المكتب الرئيسي : الرياض ص.ب ٧٧٧٠ رمز بريدي ١١٤٧٢

الفرع : جدة ص.ب ١١٢٥٧ رمز بريدي ٢١٤٥٣

الخبر :