

REPORT OF CALIBRATION

Manufacturer : Lighthouse Worldwide Solutions
Model : Handheld 3016IAQ
Serial Number : 190544003
Certificate No : 250116APC01

Calibration of the above stated instrument has been accomplished by and is fully compliant to the methods defined by ISO 21501-4. The accuracy of the standards and equipment used in the calibration are traceable to the National Institute of Standards and Technology or have been derived from acceptable values of natural physical constants. All records of work performed are maintained by Lighthouse Worldwide Solutions EMEA Operations.

All work performed is in accordance with Lighthouse Worldwide Solutions. Reproduction of this certificate and accompanying documentation is prohibited without the expressed written permission of Lighthouse Worldwide Solutions.

Test Equipment

Manufacturer	Instrument & Model	Serial Number	Calibration Date	Calibration Due
AMPTTEK	MCA8000D	01129	September 06, 2024	September 06, 2026
FLUKE	DMM179	20280165	September 06, 2024	September 06, 2026
Lighthouse	Solair Transfer Standard	131099001	June 06, 2024	June 06, 2025
TSI	4043E	40430951003	June 03, 2024	June 03, 2026

Calibration was performed under the following controlled conditions:

Temperature : 21.5 °C Relative Humidity : 44.7 %

This certifies the above named instrument conforms to the original specifications in effect at date of manufacture and test.

Threshold Voltage Settings

Channel	Size (µm)	Threshold (mV), As Found	Threshold (mV), As Left	Particle Size (nm)	Lot No.	Manufacturer	Exp. Date
1	0.30	92	95	303±6	257546	Thermo Scientific	August 31, 2025
2	0.50	643	648	510±7	255494	Thermo Scientific	June 30, 2025
3	1.00	1617	1624	1025±18	257321	Thermo Scientific	August 31, 2025
4	2.50	2561	2633	2514±27	258621	Thermo Scientific	September 30, 2025
5	5.00	5368	5464	4900±80	263147	Thermo Scientific	January 31, 2026
6	10.00	6000	6753	10000±200	265823	Thermo Scientific	April 30, 2026

Flow Rate:	Measured Flow:	2.83 LPM	(limit ±5% of nominal)	PASS
False Count:	Observed Cts:	0 p/m ³		
50% Counting Efficiency:	Size 0.303 µm:	53.79 %	(limit 30% - 70%)	PASS
100% Counting Efficiency:	Size 0.51 µm:	96.66 %	(limit 90% - 110%)	PASS
Size Resolution:	Size 0.401 µm:	8.49 %	(limit 15%)	PASS

A statement of conformity is given for Counting Efficiency (7.2), Size Resolution (7.3), False Count (7.4), Flow Rate Error (7.6.) Tests defined in the ISO 21501-4: 2018 calibration standard. The decision rule has been determined according to the ranges specified in the relevant clauses of the ISO 21501-4 standard.

Final Test Date : January 16, 2025

Signature:

Quality Assurance

Certification Date: January 16, 2025

Next calibration on this instrument is due: **January 16, 2026**

LIGHTHOUSE WORLDWIDE SOLUTIONS EMEA OPERATIONS

LF16/R04/17.10.2020

Mimar Sinan Mh, Cavusbasi Cd, Ozge Sk, Zin D Business Center No:1/26 34782 CEKMEKOY ISTANBUL TURKEY
T: +90 (216) 640 05 97 F: +90 (216) 640 05 98 E: emea@golighthouse.com W: www.golighthouse.com.tr