


RESTECH**For Trade Est.****Prop.: Eng. Ahmed Obeid****C.R.: 1010070990****C.R.: 7002028327****مؤسسة****البحاثة للتجارة****لصاحبها : المهندس أحمد عبيد****سجل تجاري ١٠١٠٠٧٠٩٩٠****سجل تجاري ٧٠٠٢٠٢٨٣٢٧****CERTIFICATE OF CONFORMITY****Certificate#****S5015****Date: 18 December 2024**

Ensure Hygiene Monitoring systems are manufactured to ISO 9000 Quality

Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive

73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by

93/68 EEC.

Ensure**Portable Luminometer**

The Ensure Plus instrument with serial number **57777** has been tested and

Performs in accordance with the product quality control specifications.

CALCHECKL LED:

Measured value(RLU)	Result
274	PASS

Engineer Signature:**Valid to: 17 DECEMBER 2025****فاكس: ٤٦٤٣٧٧٠ (١١) ٠٠٦٩٩****تلفون: ٤٦٥٣٤٨٤ (١١) ٠٠٩٦٦ / ٤٦٥٩٥٩٧ / ٤٦٥٤٠٣٣ / ٤٦٢١٦٣٣****المكتب الرئيسي: الرياض ص.ب ٧٧٠ رمز بريدي ١١٤٧٢****فاكس: ٦٧٥٧٨٤٩ (١٢) ٠٠٦٩٩****تلفون: ٦٧٣٥١٩٢ (١٢) ٠٠٩٦٦ / ٦٧٥٧٩٦٥ / ٦٧٥٧٩٦٤****الفروع : جدة ص.ب ١١٢٥٧ رمز بريدي ٢١٤٥٣****فاكس: ٨٦٧٠١٦١ (١٣) ٠٠٦٩٩****تلفون: ٨٦٧١٣١٦ (١٣) ٠٠٩٦٦****الخبر :****Head Office: P.O.Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail : restech@restech.com.sa****Branches: P.O.Box 11257 Jeddah 21453 Phone : +966 (12) 6735192 / 6757964 / 6757965 Fax: +966 (12) 6757849 E-mail: jeddah.restech@restech.com.sa****Al-Khobar :****Phone : +966 (13) 8671316 Fax : +966 (13) 8670161****E-mail: dammam.restech@restech.com.sa****www.restech.com.sa**