

Certificate # : 202513783

Issue Date : 29-Jul-25

Asset Number : GCS054028

CUSTOMER INFORMATION

Customer Name : Homeland Security Trading Est.
Address : Al Murabaa, Riyadh , K.S.A

UUT INFORMATION

Description	: Temperature Humidity Meter	Received Date	: 17-Jul-25
Manufacturer	: Fluke	Serial Number	: 63320943
Model	: 971	ID Number	: N/A
Range	: N/A	UUT Location	: N/A

CALIBRATION INFORMATION

Calibration Date	: 22-Jul-25	Result Type	: As Found
Due Date	: 22-Jul-26	Amb. Temperature	: 23 ± 5°C
Procedure Used	: Temperature and Humidity Measuring Instruments, CP/TMP/06	Relative Humidity	: 20 to 60 %RH
		Cal Site	: GCA - Riyadh

The Laboratory Quality Management system complies with the requirements of ISO/IEC 17025:2017.

TRACEABILITY INFORMATION

Measurements are traceable to the SI Units (International System of Units) via national or international laboratories. The estimated expanded uncertainty (if specified) of measurement to be associated with the results is calculated at a confidence level of approximately 95% with a coverage factor of $k = 2$.

REMARKS / DEVIATIONS IF ANY

1. The measurement results are given on the attached sheet(s). The page numbers may be numbered separately.
2. This calibration certificate applies only to the item identified and shall not be reproduced except in full without the approval by GCA.
3. The described instrument has been calibrated under the ambient conditions stated above.
4. UUT meets manufacturer's specification.
5. The UUT was received in good condition.
6. Pass/Fail decisions "if any" are based on acceptance limits chosen based on simple acceptance ($w=0$, $AL= TL$).

Standards Used				
Asset #	Description	Due Date	Certificate #	Traceability
100300698	Thermohygrometer	20-Oct-25	202417063	GCS RD
100300731	Reference Thermometer	25-Nov-25	202420334	GCS RD
100300732	RTD Probe	30-Nov-25	202420335	GCS RD

Asset # of UUT	Authorized By	 شركة المختبر الخليجي للتحقق والمطابقة GCC Lab Conformity Assessment Company خدمة المستفيدين Customer Service
 GCS054028	 Eng. Abdullah Saeed	

Page 2 of 2

Certificate #		202513783		Issue Date : 29-Jul-25					
<u>NOMINAL READING</u>		<u>UUT READING</u>		<u>ERROR</u>		<u>TOLERANCE (±)</u>		<u>EXPANDED UNCERTAINTY (±)</u>	
<u>Temperature Test</u>									
20.171	°C	20.4	°C	0.229	°C	0.5	°C	0.26	°C
24.889	°C	24.9	°C	0.011	°C	0.5	°C	0.26	°C
29.715	°C	29.5	°C	-0.215	°C	0.5	°C	0.26	°C
<u>Humidity Test</u>									
30.4	%RH	31.2	%RH	0.8	%RH	2.5	%RH	0.92	%RH
50.4	%RH	49.8	%RH	-0.6	%RH	2.5	%RH	1.2	%RH
70.4	%RH	68.0	%RH	-2.4	%RH	2.5	%RH	1.5	%RH

*Specified Limits are derived from the manufacturer specification.

*User must verify the acceptability of UUT for the intended use.

***** End of Results *****