



Calibration certificate for Audiometer

Customer: Homeland Security Trading Est.

Certificate Number
1178

Manufacturer of Audiometer: eMoyoDotNetZa

	Make	Model	Serial Number
Audiometer:	Kuduwave	Pro Tympanometer	
Earphone:	Insert	Kuduwave	0901-04770
Bone Vibrator:	BHM	BC-1 (10)	2200170174

Calibration Instrumentation

	Make	Serial Number	Due Date
Sound Level Meter:	Larson-Davis 824	A4541	2026-01-30
Preamp:		5450	2026-01-30
Calibrator:		5110	2026-01-30
Mastoid:		5677	2026-01-30
Mic Mastoid:		0169	2026-01-30
Mic 2cc coupler:		303042	2026-01-30

Environmental parameters at time of calibration

Temperature:	°C	20
Humidity:	%	22
Barometric pressure:	hPa	1022

Terms and Conditions

Coure Calibrations certifies that the air, masking, and speech conduction calibration of this audiometer conforms to the minimum specifications described in the American National Standard Institute Specifications for Audiometers (ANSI S3.6:2018 R2023), IEC 60645-1:2017 and the South African Bureau of Standards (SANS 10154-1).

Calibration data for this calibration is supplied with this certificate (Upon Request).

This certificate becomes invalid if this audiometer is misused / roughly handled or broken/subjected to repairs / moved by road, air, or rail without an anti-vibration case and if no subjective calibration test has been done before testing commences at a new site.

This calibration certificate is issued under the ISO 13485 certification license of GeoAxon Global (Pty) Ltd



2025-03-11

2026-03-11

Technician

Calibration Date

Next Calibration Date

Scientist

Mr F Simutowe

Dr.H L Koekemoer
(MBChB)