

**CERTIFICATE OF CONFORMITY****Certificate#****M-3226****Date: 10 July 2024**

Ensure Hygiene Monitoring systems are manufactured to ISO 9000 Quality Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive 73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by 93/68 EEC.

Ensure**Portable Luminometer**

The Ensure Plus instrument with serial number **092887** has been tested and Performs in accordance with the product quality control specifications.

ATP Positive Control:

Positive Rod	Measured value (RLU)	Result
140-220	162	PASS

ATP Negative Control:

Negative Rod	Measured value (RLU)	Result
0 - 4	0	PASS

Engineer Signature:**Valid until: 10 July 2025**