

**CERTIFICATE OF CONFORMITY****Certificate#****M-3227****Date: 10 July 2024**

System Sure Monitoring systems are manufactured to ISO 9000 Quality

Standard and comply with IEC1010-1 and EN61010-1:2001.

The system satisfies the requirements of, article 11 of the Low Voltage Directive

73/23/EEC the EMC Directive(s) 89/336/EEC and 92/31/EEC as realigned by

93/68 EEC.

SystemSure Plus**Portable Luminometer**

SystemSure Plus instrument with serial number **072399** has been tested and

Performs in accordance with the product quality control specifications.

ATP Positive Control:

Positive Rod	Measured value (RLU)	Result
140-220	159	PASS

ATP Negative Control:

Negative Rod	Measured value (RLU)	Result
0 - 4	0	PASS

Engineer Signature:



Valid until: 10 July 2025

فاكس : ٤٦٤٣٧٧٠ (١١) ٠٠٩٦٦

فاكس : ٦٧٥٧٨٤٩ (١٢) ٠٠٩٦٦

فاكس : ٨٦٧٠١٦١ (١٣) ٠٠٩٦٦

تلفون : ٤٦٥٤٨٤ (١١) ٠٠٩٦٦ / ٤٦٥٩٥٩٧ / ٤٦٢١٦٣٣

تلفون : ٦٧٣٥١٩٢ (١٢) ٠٠٩٦٦ / ٦٧٥٧٩٦٤ / ٦٧٥٧٩٦٥

تلفون : ٨٦٧١٣١٦ (١٣) ٠٠٩٦٦

المكتب الرئيسي : الرياض ص.ب ٧٧٠ رمز بريدي ١١٤٧٢

الفروع : جدة ص.ب ١١٢٥٧ رمز بريدي ٢١٤٥٣

الخبر :

Head Office: P.O. Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail: restech@restech.com.sa

Branches: P.O. Box 11257 Jeddah 21453 Phone : +966 (12) 6735192 / 6757964 / 6757965 Fax: +966 (12) 6757849 E-mail: jeddah.restech@restech.com.sa

Al-Khobar : Phone : +966 (13) 8671316 Fax : +966 (13) 8670161

E-mail: dammam.restech@restech.com.sa

www.restech.com.sa