


RESTECH**For Trade Est.****Prop.: Eng. Ahmed Obeid****C.R.: 1010070990****C.R.: 7002028327**

مؤسسة

البحاثة للتجارة

إصاحبها : المهندس أحمد عبيد

سجل تجاري ١٠١٠٧٠٩٩٠

سجل تجاري ٧٠٠٢٠٢٨٣٢٧

Instrument Quality Certificate**INSTRUMENT :****SN:****SOFTWARE VERSION :****HI991001****TA02310050****1.03**

Hanna Instruments certifies that this instrument has been produced, calibrated and tested to meet all applicable Hanna Instruments procedures, using standards and reference instruments, the accuracy of which is traceable to the National Institute of Standards (NIST) in the USA or to internationally acceptable national physical standards. The standards and reference instruments used in calibration and testing are supported by a calibration system which meets requirements of ISO 9001. The following tests have been performed according with the reference from the QC Procedure of the meter.

THE RESULTS ARE LISTED BELOW

A. Functionality tests	REFERENCE	RESULT
A.1.Switch On/offtest	8.3.	PASSED
A.2.Keys and LCD test	8.3.	PASSED
A3.Battery level test	8.3.	PASSED
A.4. Buzzer Sound test	8.3.	PASSED
A.5. Voltage and Consumption test	8.3.	PASSED
A.6.Default Settings test	8.3.	PASSED
A.7.Probe Connectivity test	8.4.	PASSED
A.8.Calibration test (pH)	8.4.	PASSED
A.9. Measurements test (pH, Temperature)	8.4.	PASSED

B. Aesthetic Control	REFERENCE	RESULT
B.1. Visual Inspection	B.1.	PASSED
B.2. Labeling and Marking	B.1.	PASSED

Calibration, functionality test, aesthetic control and packing have been

DATE :18/5/2025

INSPECTOR: HISHAM A.MONIEM SULIEMAN

VALID TILL : 17/5/2026



فاكس: ٤٦٤٣٧٧٠ (١١) ٠٠٦٩٩

فاكس: ٦٧٥٧٨٤٩ (١٢) ٠٠٦٩٩

فاكس: ٨٦٧٠١٦١ (١٣) ٠٠٦٩٩

تلفون: ٤٦٥٣٤٨٤ (١١) ٠٠٩٦٦ / ٤٦٥٩٥٩٧ / ٤٦٥٤٠٣٣ / ٤٦٢١٦٣٣

تلفون: ٦٧٣٥١٩٢ (١٢) ٠٠٩٦٦ / ٦٧٥٧٩٦٥ / ٦٧٥٧٩٦٤

تلفون: ٨٦٧١٣١٦ (١٣) ٠٠٩٦٦

المكتب الرئيسي: الرياض ص.ب ٧٧٧٠ رمز بريدي ١١٤٧٢

الفرع: جدة ص.ب ١١٢٥٧ رمز بريدي ٢١٤٥٣

الخبر:

Head Office: P.O.Box 7770 Riyadh 11472 Phone : +966 (11) 4653484 / 4659597 / 4654033 / 4621633 Fax : +966 (11) 4643770 E-mail : restech@restech.com.sa

Branches: P.O.Box 11257 Jeddah 21452 Phone : +966 (12) 6735183 / 6757004 / 6757005 Fax : +966 (12) 6757010 E-mail : restech@restech.com.sa